

SOLUTION CHANGE

DataMaster DMT: 100169
Location: Bellows Falls PD
Date: 08/08/2016
Performed by: MICHAEL J KEEFE



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-24-100
Range = 0.094 - 0.104
Average = 0.101 g/210L
Std Dev = 0.0000

RCD
8/8/16

Simulator Temperature: 34.1°C

Performed by

Michael J Keefe

Date

08/08/2016